

CERCELL A/S  
Nørrelundvej 10  
2730 Herlev  
Denmark

**Processing Facility** Denmark E-Beam  
**Sales Order #** 2777932  
**P.O.#** 1868  
**Processing Start** 24/5/2021 04:11:44 GMT  
**Processing End** 24/5/2021 04:50:03 GMT

Work Order # 2892379		Customer Requested Target Dose		4 x 8 kGy - 4 sides	
Qty	UOM	Item Number	Item Description	Customer Lot Number	Customer Load Number
12	CR	CC_CAD35	CAD35	96	Palle #1
12	CR	CC_CAD35	CAD35	96	Palle #2
6	CR	CC_CAD35	CAD35	96	Palle #3
7	CR	CC_CAD12C	CAD12	96	Palle #4
5	CR	CC_CAD21-1	CAD21	96	Palle #4
42	CR	<b>Total</b>			


**Quality Test Summary**

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result
650.00	Reference Minimum Dose	7.2 kGy	8.8 kGy	7.8 KGY
650.00	Reference Maximum Dose	7.2 kGy	8.8 kGy	8.0 KGY

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

All boxes including the ones with item number CC\_Merck13L have been flipped and irradiated from all 4 sides with a target dose of 8.0kGy each time, to an accumulated target dose of 32kGy.

Certified By:



Date:

25 MAY 2021