



Certificate of Processing

R55480106

10/03/2014 GMT

CERCELL APS
MALMMOSEVEJ, 19C
2840 HOLTE
Denmark

Processing Facility Denmark E-Beam
Sales Order # 1100084
P.O.# Rekv. 30
Processing Date 9/3/2014 GMT

Work Order # 1256984


Material Test

Qty	UOM	Item Number	Item Description	Customer Lot Number	Customer Load Number
5	CR		Manual Processing	CCLVARIOUS/Batch nr. 10	2x16 kGy
3	CR		Manual Processing	CCLVARIOUS/Batch nr. 10	2x16 kGy
1	CR		Manual Processing	CCLVARIOUS/Batch nr. 10	2x16 kGy
9	CR		Total		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result
650.00	Reference Minimum Dose	.1 kGy	82.8 kGy	15.8 KGY
650.00	Reference Maximum Dose	.1 kGy	82.8 kGy	16.4 KGY

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Certified By:  Date: 10 Mar 2014